

FUJI DRI-CHEM CALIBRATOR CP (CRP)

Intended use

FUJI DRI-CHEM CALIBRATOR CP (CRP) is the standard solution specially designed for FUJI DRI-CHEM system. This product is used to calibrate C-reactive protein test using FUJI DRI-CHEM SLIDE CRP-S.

For *in vitro* diagnostic use only.

Reagents

This product is an aqueous solution prepared from pooled human serum. It is added with C-reactive protein purified from human serum and other ingredients.

[Reactive ingredients]

Human C-reactive protein	CP-1	9 mg/L
	CP-2	30 mg/L
	CP-3	71 mg/L

[Other ingredients]

- Glucose
- Sodium azide 0.1 %

Test object

Human C-reactive protein

System applied

1. Analyzer: FUJI DRI-CHEM ANALYZER
2. Slide: FUJI DRI-CHEM SLIDE CRP-S
3. Diluent: FUJI DRI-CHEM DILUENT DL (CRP)

Procedure

1. Wear proper gloves, glasses and other protective gear for your safety.
2. This product should be stored between 2–8 °C (35.6–46.4 °F) and warmed up to room temperature before use.
3. Dilute this calibrator 21 times by FUJI DRI-CHEM DILUENT DL (CRP). This can be done automatically in the analyzer equipped with auto dilution function.
4. Select calibration mode on the analyzer and assay this product in the same way as patient specimens.
For further details, consult "INSTRUCTION MANUAL" for FUJI DRI-CHEM analyzer.

Assigned value

	CP-1	CP-2	CP-3	Unit
CRP	9	30	71	mg/L
	0.9	3.0	7.1	mg/dL

Warnings and precautions

1. FUJI DRI-CHEM CALIBRATOR CP (CRP) has been found to be non-reactive for HBs Ag (hepatitis-B virus antigen), anti-HCV (hepatitis-C virus) and anti-HIV (human immunodeficiency virus) antibodies. However, as there is no absolute proof of non-infectiousness, this product should be handled in the same way as patient specimens.
2. Please do not remove holders attached to the bottles.
3. When the cap is opened, a liquid membrane produced by surface tension is often observed at the mouth of the bottle. Be sure to remove it before sampling because it may interfere with normal pipetting.

4. For each calibration, please do not keep bottles open over five minutes to avoid a concentration of the liquid. And please do not use the liquid for calibration more than ten times even within an expiry.
5. Seal up tightly after use to avoid concentration.
6. If a volume of the liquid is decreased by spillages etc, please refrain from using the liquid for calibration, because it might be to be concentrated.
7. Since this product contains 0.1 % of sodium azide as a preservative, some special precautions are required.
 - In case of ingestion, eye or skin contact, wash with plenty of water immediately and consult medical personnel.
8. The residual liquid is categorized as infectious waste. Make sure to dispose it in accordance with the Waste Disposal Law and other related regulations, which prescribe the proper method of disposal, such as incineration, sterilization or disinfection.
9. This product cannot be applied to other devices not specified by the manufacturer.



[Xn, Harmful, R: 22 S: 23 - Contains sodium azide]

- Harmful if swallowed. Do not breathe gas/fumes/vapour/spray.
- Gesundheitsschädlich beim Verschlucken. Gas/Rauch/Dampf/Aerosol nicht einatmen.
- Nocif en cas d'ingestion. Ne pas respirer les gaz/fumées/vapeurs/aérosols.
- Nocivo per ingestione. Non respirare i gas/fumi/vapori/aerosoli.
- Nocivo por ingestión. No respirar los gases/humos/vapores/aerosoles.
- Nocivo por ingestão. Não respirar os gases/vapores/fumos/aerossóis.
- Farlig ved indtagelse. Undgå indånding af gas/røg/dampe/aerosol-løser.
- Terveydelle haitallista nieltynä. Välttäävä kaasin/huuron/höyryn/sumun hengittämistä.
- Επιβλαβές σε περίπτωση κατάποσής, Μην αναπνέετε αέρια/αεραθμίσεις/ατμούς/εκνεφώματα.
- Schadelijk bij opname door de mond. Gas/rook/damp/sputnevel niet inademen.
- Farlig ved svelgning. Unngå indånding av gass/røyk/damp/sprøyteåke.
- Farligt vid förtäring. Undvik inandning av gas/rök/ånga/dimma.
- Zdravi škodlivý při požití. Nevdechujte plyn/dým/páry/aerosoly.
- Kahjulik allaneelamisel. Vältida gaasi/suitsu/auru/udu sissehingamist.
- Lenyelve ártalmas. A keletkező gáz/füstöl/gőzt/permetet nem szabad belélegezni.
- Kaitgits norjot. Izvältitries no gāzes vai dūmu, vai tvaiku, vai aerosolu ieeļpošanas.
- Kenksminga priarijus. Neikvėpti dūjų dūmų garų aerosolius.
- Dział szkodliwe po połknięciu. Nie wdychać gazu/dymu/pary/rozpylonej cieczy.
- Škodlivý po požití. Nevdychujte plyn/dým/pary/aerosoly.
- Zdravju škodljivo pri zaužitju. Ne vdihavati plina/dima/hlapov/meglice.
- Вредный при проглатывании. Не вдыхайте газ/дым/пар/распыления.
- Yutulduğunda, zararlıdır. Gaz/duman/buhar/püskürtüyü solumayınız.

Storage and shelf life

1. Storage condition: Between 2–8 °C (35.6–46.4 °F)
2. Expiry date: Indicated on the label of the vial and the carton box.

Contents

- CP-1 1 mL x 1 bottle : CP-2 1 mL x 1 bottle
- CP-3 1 mL x 1 bottle : Instructions for Use 1



<http://www.fujifilm.com/products/medical/>



FUJIFILM Europe GmbH
Heesenstr. 31, D-40549 Düsseldorf, GERMANY



FUJIFILM Corporation
26-30, Nishiazabu 2-chome, Minato-ku Tokyo, JAPAN

